

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001149  
10:10 30/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:56  
PRESSURIZING 3:02  
STERILIZING 4:06  
136.0 C 220kPa 4:06  
136.4 C 225kPa 4:36  
136.4 C 223kPa 5:06  
136.4 C 225kPa 5:36  
136.4 C 223kPa 6:06  
136.1 C 223kPa 6:36  
136.0 C 223kPa 7:06  
136.4 C 224kPa 7:36  
136.4 C 224kPa 7:36  
VENTING 7:36  
AIR DRYING 8:25  
CYCLE COMPLETE 12:54  
-----

Digital Signature #  
3296E9F27830F7BC  
-----