
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001979
13:00 26/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:42
STERILIZING 3:45
136.5 C 227kPa 3:45
136.3 C 225kPa 4:15
136.4 C 226kPa 4:45
136.4 C 226kPa 5:15
136.4 C 226kPa 5:45
136.2 C 225kPa 6:15
136.3 C 222kPa 6:45
136.3 C 225kPa 7:15
136.3 C 225kPa 7:15
VENTING 7:15
AIR DRYING 8:09
CYCLE COMPLETE 8:16

Digital Signature #
B8DBF4C895BB2D1B
