
Midmark M11 - v1.0.0
Total Cycles: 3294

Name: RITTER 1
Sterilizer ID: V2178198
05/29/2023 08:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.4 F 0.0
2:00 199.2 F 0.0
4:00 205.6 F 2.9
6:00 210.4 F 7.8
8:00 225.5 F 12.8
10:00 230.6 F 8.6
12:00 239.7 F 10.9
14:00 243.0 F 12.1
16:00 241.4 F 10.9
18:00 254.6 F 17.5
20:00 265.9 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.8 F 28.8
2:00 271.9 F 28.1
3:00 271.7 F 28.4
4:00 272.7 F 28.7

Min 271.2 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 223.1 F 0.0
10:00 227.3 F 0.0
15:00 210.1 F 0.0
20:00 206.6 F 0.0
25:00 204.2 F 0.0
30:00 201.8 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:00:40
CYCLE COMPLETE

05/29/2023 09:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____