
Midmark M11 - v1.0.5
Total Cycles: 3728

Name: RITTER2
Sterilizer ID: V2313228
06/21/2024 01:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.2 F 0.0
2:00 124.4 F 1.4
4:00 150.2 F 4.0
6:00 175.1 F 7.2
8:00 192.8 F 11.2
10:00 211.3 F 4.1
12:00 215.7 F 8.5
14:00 227.8 F 13.0
16:00 230.8 F 9.4
18:00 237.6 F 10.0
20:00 245.1 F 13.3
22:00 256.8 F 19.4
24:00 266.9 F 25.9
26:00 273.1 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.6 F 29.9
2:00 271.7 F 29.8
3:00 272.1 F 29.7
4:00 271.6 F 29.8

Min 271.3 F 29.3
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.5 F 0.0
5:00 225.6 F 0.0
10:00 221.3 F 0.0
15:00 203.3 F 0.0
20:00 200.5 F 0.0
25:00 207.6 F 0.0
30:00 209.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 26:18
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:04:56
CYCLE COMPLETE

06/21/2024 02:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____