
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 001438
12:47 09/16/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:58
PRESSURIZING 5:57
STERILIZING 7:28
136.4 C 220kPa 7:28
136.4 C 221kPa 7:58
136.5 C 223kPa 8:28
136.2 C 219kPa 8:58
136.4 C 222kPa 9:28
136.5 C 223kPa 9:58
136.2 C 221kPa 10:28
136.4 C 221kPa 10:58
VENTING 10:58
AIR DRYING 11:59
CYCLE COMPLETE 41:53

Digital Signature #
3503BD97A7B051F7
