
Midmark M11 - v1.0.0
Total Cycles: 2515

Name: RITTER 1
Sterilizer ID: V2178198
11/04/2022 09:41 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.4 F 0.0
2:00 160.6 F 2.8
4:00 174.6 F 6.9
6:00 198.8 F 12.3
8:00 220.2 F 6.4
10:00 237.2 F 12.9
12:00 239.1 F 10.5
14:00 237.2 F 8.9
16:00 252.9 F 16.5
18:00 265.6 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.1 F 28.2
2:00 272.9 F 28.8
3:00 271.8 F 28.3

Min 271.4 F 27.8
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 229.4 F 0.0
Dry Cycle Aborted

FILLING: 1:08
HEATING: 19:34
STERILIZING: 3:00
VENTING: 2:47
DRYING: 3:09
TOTAL CYCLE: 00:29:38

11/04/2022 10:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____