
Midmark M11 - v1.0.5
Total Cycles: 4352

Name: RITTER2
Sterilizer ID: V2313228
02/12/2025 11:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.7 F 0.0
2:00 119.5 F 0.9
4:00 121.5 F 2.9
6:00 144.9 F 6.2
8:00 180.2 F 9.7
10:00 196.8 F 9.1
12:00 214.7 F 6.3
14:00 226.6 F 10.3
16:00 233.8 F 8.6
18:00 238.6 F 11.5
20:00 236.7 F 9.4
22:00 248.9 F 15.3
24:00 259.9 F 21.4
26:00 269.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.7
1:00 272.0 F 29.6
2:00 272.1 F 29.6
3:00 272.3 F 29.6
4:00 272.5 F 29.8

Min 271.9 F 29.6
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 210.3 F 0.0
10:00 206.9 F 0.0
15:00 211.8 F 0.0
20:00 188.3 F 0.0
25:00 172.3 F 0.0
30:00 168.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 27:49
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:06:24
CYCLE COMPLETE

02/12/2025 12:58 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____