
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004675
15:55 07/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:03
STERILIZING 4:03
136.4 C 222kPa 4:03
136.3 C 223kPa 4:33
136.4 C 223kPa 5:03
136.3 C 223kPa 5:33
136.2 C 222kPa 6:03
136.3 C 224kPa 6:33
136.4 C 225kPa 7:03
136.1 C 222kPa 7:33
136.1 C 222kPa 7:33
VENTING 7:33
AIR DRYING 8:28
CYCLE COMPLETE 9:55

Digital Signature #
FF234C225446785A
