

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4286

Name: RITTER 1  
Sterilizer ID: V2178198  
02/08/2024 03:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.7 F 0.0  
2:00 131.2 F 1.3  
4:00 133.8 F 3.8  
6:00 156.0 F 7.0  
8:00 179.5 F 10.9  
10:00 217.6 F 4.4  
12:00 219.1 F 7.4  
14:00 228.7 F 11.7  
16:00 230.5 F 7.6  
18:00 242.3 F 12.9  
20:00 240.6 F 10.9  
22:00 252.5 F 16.7  
24:00 262.8 F 22.8  
26:00 272.7 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.0  
1:00 272.2 F 29.1  
2:00 272.6 F 29.4  
3:00 272.3 F 29.6  
4:00 271.8 F 28.9

Min 271.7 F 28.9  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
5:00 222.3 F 0.0  
10:00 233.8 F 0.0  
15:00 229.9 F 0.0  
20:00 232.8 F 0.0  
25:00 231.1 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 27:10  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:05:53  
CYCLE COMPLETE

02/08/2024 04:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_