
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002271
13:34 15/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:56
STERILIZING 3:05
136.4 C 226kPa 3:05
136.4 C 226kPa 3:35
136.3 C 224kPa 4:05
136.0 C 223kPa 4:35
136.4 C 225kPa 5:05
136.4 C 225kPa 5:35
136.4 C 226kPa 6:05
136.4 C 226kPa 6:35
136.4 C 226kPa 6:35
VENTING 6:35
AIR DRYING 7:27
CYCLE COMPLETE 37:27

Digital Signature #
1CB576647D1EDDE8
