
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000383
12:08 22/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:57
STERILIZING 4:14
136.3 C 223kPa 4:14
136.2 C 230kPa 4:44
136.1 C 227kPa 5:14
136.3 C 232kPa 5:44
135.9 C 228kPa 6:14
136.4 C 232kPa 6:44
136.4 C 232kPa 7:14
136.4 C 232kPa 7:44
136.4 C 232kPa 7:44
VENTING 7:45
AIR DRYING 8:38
CYCLE COMPLETE 23:38

Digital Signature #
EBB9806793429AA9
