

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3594

Name: RITTER2  
Sterilizer ID: V2313228  
05/07/2024 08:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.3 F 0.0  
2:00 71.8 F 0.0  
4:00 90.9 F 0.0  
6:00 134.1 F 1.2  
8:00 151.2 F 4.4  
10:00 168.1 F 8.2  
12:00 188.6 F 11.6  
14:00 203.7 F 3.7  
16:00 217.0 F 7.2  
18:00 231.6 F 10.9  
20:00 234.1 F 7.2  
22:00 239.5 F 11.7  
24:00 237.1 F 9.5  
26:00 250.0 F 15.6  
28:00 260.8 F 21.8  
30:00 271.0 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.9  
1:00 273.0 F 29.6  
2:00 271.7 F 28.7  
3:00 272.7 F 29.8  
4:00 272.4 F 29.1

Min 271.6 F 28.7  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.6 F 0.0  
5:00 191.7 F 0.0  
10:00 192.9 F 0.0  
15:00 192.0 F 0.0  
20:00 190.7 F 0.0  
25:00 198.2 F 0.0  
30:00 204.4 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 30:48  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:08:45  
CYCLE COMPLETE

05/07/2024 09:40 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_