
Midmark M11 - v1.0.0
Total Cycles: 4969

Name: RITTER 1
Sterilizer ID: V2178198
10/07/2024 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.0 F 0.0
2:00 79.5 F 0.0
4:00 81.2 F 0.0
6:00 114.8 F 0.1
8:00 146.6 F 0.5
10:00 174.3 F 2.6
12:00 209.1 F 5.8
14:00 213.9 F 9.3
16:00 225.7 F 13.1
18:00 227.5 F 6.7
20:00 238.1 F 10.9
22:00 232.8 F 7.2
24:00 244.9 F 12.9
26:00 242.0 F 11.1
28:00 254.3 F 17.4
30:00 265.1 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.1
1:00 272.1 F 28.7
2:00 271.9 F 28.4
3:00 271.6 F 28.4
4:00 272.4 F 28.8

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 204.3 F 0.0
10:00 189.2 F 0.0
15:00 190.2 F 0.0
20:00 211.4 F 0.0
25:00 200.4 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 31:45
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:09:56
CYCLE COMPLETE

10/07/2024 09:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____