
Midmark M11 - v1.0.5
Total Cycles: 47

Name: RITTER2
Sterilizer ID: V2313228
05/27/2021 07:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.8 F 0.0
2:00 75.2 F 0.2
4:00 79.3 F 1.2
6:00 105.0 F 3.3
8:00 134.5 F 5.9
10:00 161.0 F 8.9
12:00 185.6 F 12.5
14:00 201.6 F 3.8
16:00 214.7 F 7.4
18:00 227.0 F 11.5
20:00 229.7 F 7.7
22:00 240.5 F 12.8
24:00 240.8 F 11.4
26:00 252.2 F 17.4
28:00 263.1 F 23.5
30:00 272.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.7 F 29.5
2:00 272.1 F 29.2
3:00 272.1 F 29.6
4:00 272.1 F 29.2

Min 271.6 F 29.1
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.0 F 0.0
5:00 236.4 F 0.0
10:00 231.9 F 0.0
15:00 222.1 F 0.0
20:00 173.1 F 0.0
25:00 203.9 F 0.0
30:00 197.0 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 30:09
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:08:24
CYCLE COMPLETE

05/27/2021 08:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____