
Midmark M11 - v1.0.5
Total Cycles: 3996

Name: RITTER2
Sterilizer ID: V2313228
10/03/2024 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.1 F 0.0
2:00 138.0 F 0.8
4:00 158.3 F 3.9
6:00 183.3 F 7.5
8:00 202.2 F 11.4
10:00 217.0 F 4.5
12:00 229.2 F 8.3
14:00 239.3 F 12.5
16:00 236.0 F 9.2
18:00 240.4 F 10.9
20:00 245.6 F 13.2
22:00 256.5 F 19.1
24:00 266.5 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.6 F 29.2
2:00 271.3 F 29.1
3:00 272.5 F 29.1
4:00 271.1 F 28.9

Min 271.1 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 217.2 F 0.0
10:00 206.3 F 0.0
15:00 203.8 F 0.0
20:00 205.6 F 0.0
25:00 206.5 F 0.0
30:00 212.7 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 25:41
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:04:39
CYCLE COMPLETE

10/03/2024 04:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____