
Midmark M11 - v1.0.5
Total Cycles: 740

Name: RITTER2
Sterilizer ID: V2313228
01/31/2022 05:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.0 F 0.0
2:00 183.5 F 3.7
4:00 206.7 F 9.5
6:00 220.4 F 6.9
8:00 235.3 F 11.1
10:00 237.6 F 9.8
12:00 237.6 F 9.2
14:00 253.1 F 17.0
16:00 265.6 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.7 F 29.3
2:00 271.8 F 28.5
3:00 271.9 F 29.1
4:00 272.7 F 29.3

Min 271.2 F 28.3
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 237.8 F 0.0
10:00 219.7 F 0.0
15:00 210.2 F 0.0
20:00 224.2 F 0.0
25:00 229.8 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 2:28
HEATING: 17:39
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:14
CYCLE COMPLETE

01/31/2022 06:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____