
Midmark M11 - v1.0.5
Total Cycles: 3791

Name: RITTER2
Sterilizer ID: V2313228
07/16/2024 05:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.9 F 0.0
2:00 147.8 F 2.1
4:00 162.2 F 5.4
6:00 185.6 F 9.8
8:00 212.5 F 7.5
10:00 223.4 F 8.8
12:00 233.4 F 12.3
14:00 236.8 F 11.2
16:00 236.3 F 9.6
18:00 249.4 F 16.0
20:00 260.9 F 22.5
22:00 271.3 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 30.0
1:00 272.4 F 30.0
2:00 272.3 F 30.1
3:00 272.4 F 30.1
4:00 272.6 F 30.3

Min 271.6 F 29.9
Max 273.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 220.8 F 0.0
10:00 219.4 F 0.0
15:00 211.9 F 0.0
20:00 199.7 F 0.0
25:00 215.3 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 23:13
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:02:19
CYCLE COMPLETE

07/16/2024 06:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____