

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4815

Name: RITTER 1  
Sterilizer ID: V2178198  
08/14/2024 11:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 121.5 F 0.0  
2:00 122.6 F 0.5  
4:00 144.1 F 1.0  
6:00 172.1 F 2.3  
8:00 192.1 F 5.3  
10:00 212.2 F 8.3  
12:00 219.7 F 12.4  
14:00 222.8 F 6.1  
16:00 232.5 F 10.5  
18:00 234.0 F 8.3  
20:00 242.3 F 12.1  
22:00 239.7 F 10.2  
24:00 251.7 F 16.1  
26:00 261.9 F 22.0  
28:00 271.2 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.2  
1:00 271.6 F 28.6  
2:00 272.3 F 28.9  
3:00 273.1 F 29.4  
4:00 272.8 F 29.1

Min 271.0 F 28.1  
Max 273.2 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.0 F 0.0  
5:00 205.8 F 0.0  
10:00 210.5 F 0.0  
15:00 208.9 F 0.0  
20:00 218.7 F 0.0  
25:00 226.5 F 0.0  
30:00 226.2 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 28:52  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:07:32  
CYCLE COMPLETE

08/14/2024 12:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_