
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000521
12:10 28/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:15
STERILIZING 4:25
136.4 C 221kPa 4:25
136.2 C 232kPa 4:55
136.2 C 230kPa 5:25
136.2 C 232kPa 5:55
136.4 C 233kPa 6:25
136.4 C 232kPa 6:55
136.3 C 231kPa 7:25
136.4 C 231kPa 7:55
136.4 C 231kPa 7:55
VENTING 7:55
AIR DRYING 8:49
CYCLE COMPLETE 11:58

Digital Signature #
6D967CCA6C41514C
