
Midmark M11 - v1.0.0
Total Cycles: 3241

Name: RITTER 1
Sterilizer ID: V2178198
05/16/2023 11:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.5 F 0.0
2:00 198.0 F 2.6
4:00 210.3 F 7.8
6:00 225.1 F 11.5
8:00 232.6 F 9.3
10:00 239.7 F 9.4
12:00 244.7 F 13.0
14:00 243.2 F 11.9
16:00 256.1 F 18.7
18:00 267.4 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.3 F 29.2
2:00 271.9 F 28.7
3:00 271.5 F 28.8
4:00 272.4 F 29.1

Min 271.2 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 231.6 F 0.0
10:00 235.1 F 0.0
15:00 230.3 F 0.0
20:00 229.9 F 0.0
25:00 224.3 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

05/16/2023 12:29 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____