
Midmark M11 - v1.0.5
Total Cycles: 162

Name: RITTER2
Sterilizer ID: V2313228
07/05/2021 01:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.6 F 0.0
2:00 186.2 F 3.7
4:00 191.7 F 8.4
6:00 211.9 F 12.7
8:00 223.6 F 8.8
10:00 235.9 F 11.3
12:00 240.4 F 12.2
14:00 241.5 F 11.4
16:00 254.0 F 18.0
18:00 265.3 F 24.8
20:00 271.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.7 F 29.7
2:00 272.5 F 30.1
3:00 271.6 F 29.3
4:00 272.5 F 29.9

Min 271.5 F 29.3
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 194.3 F 0.0
10:00 184.1 F 0.0
15:00 175.9 F 0.0
20:00 165.1 F 0.0
25:00 165.0 F 0.0
Dry Cycle Aborted

FILLING: 2:07
HEATING: 20:31
STERILIZING: 4:00
VENTING: 3:12
DRYING: 29:24
TOTAL CYCLE: 00:59:14

07/05/2021 02:33 PM
ACCEPTED: X
REJECTED:
Operator: U01