

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4320

Name: RITTER 1  
Sterilizer ID: V2178198  
02/22/2024 09:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.5 F 0.0  
2:00 73.6 F 0.2  
4:00 79.5 F 1.1  
6:00 100.9 F 3.1  
8:00 133.1 F 5.9  
10:00 160.6 F 9.4  
12:00 183.7 F 13.2  
14:00 205.0 F 4.1  
16:00 215.8 F 7.3  
18:00 226.9 F 10.9  
20:00 229.1 F 7.1  
22:00 238.7 F 11.0  
24:00 233.9 F 8.0  
26:00 247.4 F 14.1  
28:00 258.0 F 19.9  
30:00 268.0 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.3 F 29.0  
2:00 271.2 F 28.4  
3:00 272.4 F 29.1  
4:00 271.4 F 28.4

Min 271.2 F 28.3  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 208.8 F 0.0  
10:00 212.2 F 0.0  
15:00 189.7 F 0.0  
20:00 218.9 F 0.0  
25:00 232.1 F 0.0  
30:00 226.4 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 31:24  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:09:41  
CYCLE COMPLETE

02/22/2024 10:55 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_