
Midmark M11 - v1.0.0
Total Cycles: 1290

Name: RITTER 1
Sterilizer ID: V2178198
10/22/2021 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.5 F 0.0
2:00 181.1 F 3.4
4:00 198.8 F 8.2
6:00 218.1 F 11.2
8:00 231.5 F 8.4
10:00 241.5 F 12.4
12:00 241.2 F 11.3
14:00 238.7 F 9.7
16:00 253.1 F 16.7
18:00 264.8 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 29.0
2:00 271.7 F 28.1
3:00 272.1 F 28.9
4:00 272.6 F 28.8

Min 271.3 F 28.0
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 228.8 F 0.0
10:00 210.1 F 0.0
15:00 191.1 F 0.0
20:00 186.2 F 0.0
25:00 189.6 F 0.0
30:00 213.8 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 19:54
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

10/22/2021 12:03 PM
ACCEPTED: X
REJECTED:
Operator: U02