
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002078
9:56 20/07/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:37
STERILIZING 4:39
136.0 C 220kPa 4:39
136.3 C 223kPa 5:09
136.4 C 221kPa 5:39
136.2 C 223kPa 6:09
136.4 C 224kPa 6:39
136.4 C 224kPa 7:09
136.3 C 221kPa 7:39
136.4 C 224kPa 8:09
136.4 C 224kPa 8:09
VENTING 8:09
AIR DRYING 9:04
CYCLE COMPLETE 9:09

Digital Signature #
EFA903CE85F168C1
