

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3550

Name: RITTER 1  
Sterilizer ID: V2178198  
07/19/2023 12:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.4 F 0.0  
2:00 171.7 F 3.0  
4:00 184.2 F 6.5  
6:00 207.9 F 11.7  
8:00 218.8 F 5.6  
10:00 231.7 F 10.7  
12:00 232.1 F 7.1  
14:00 244.9 F 13.5  
16:00 243.9 F 12.2  
18:00 256.5 F 18.8  
20:00 267.6 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.7 F 28.9  
2:00 271.6 F 28.4  
3:00 272.8 F 29.0  
4:00 271.8 F 28.3  
  
Min 271.6 F 28.3  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.7 F 0.0  
5:00 215.8 F 0.0  
10:00 228.1 F 0.0  
15:00 228.6 F 0.0  
20:00 227.0 F 0.0  
25:00 225.7 F 0.0  
30:00 225.8 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 21:22  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:59:44  
CYCLE COMPLETE

07/19/2023 01:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_