
Midmark M11 - v1.0.5
Total Cycles: 3882

Name: RITTER2
Sterilizer ID: V2313228
08/22/2024 08:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 70.6 F 0.0
4:00 76.3 F 0.0
6:00 105.7 F 0.9
8:00 142.2 F 3.7
10:00 168.1 F 6.6
12:00 190.4 F 9.9
14:00 221.2 F 8.1
16:00 221.9 F 7.1
18:00 234.0 F 11.1
20:00 228.4 F 6.8
22:00 241.2 F 12.1
24:00 238.3 F 9.9
26:00 250.6 F 15.8
28:00 261.2 F 21.9
30:00 271.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.1 F 29.1
2:00 272.7 F 29.3
3:00 272.8 F 29.6
4:00 272.6 F 29.8

Min 271.6 F 29.1
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 235.1 F 0.0
10:00 185.1 F 0.0
15:00 180.0 F 0.0
20:00 176.5 F 0.0
25:00 173.9 F 0.0
30:00 172.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:20
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:09:27
CYCLE COMPLETE

08/22/2024 09:36 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____