
Midmark M11 - v1.0.5
Total Cycles: 2133

Name: RITTER2
Sterilizer ID: V2313228
03/09/2023 11:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.8 F 0.0
2:00 178.7 F 3.2
4:00 194.9 F 7.9
6:00 212.1 F 12.4
8:00 226.8 F 9.2
10:00 235.4 F 9.4
12:00 243.9 F 13.2
14:00 244.2 F 12.5
16:00 257.4 F 19.7
18:00 269.3 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 271.8 F 29.3
2:00 272.4 F 29.3
3:00 271.3 F 28.6
4:00 271.5 F 29.2

Min 271.0 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 216.9 F 0.0
10:00 220.4 F 0.0
15:00 224.5 F 0.0
20:00 228.3 F 0.0
25:00 227.5 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 18:56
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:54
CYCLE COMPLETE

03/09/2023 12:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____