
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000291
11:34 16/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:59
STERILIZING 4:18
136.0 C 223kPa 4:18
136.4 C 233kPa 4:48
136.4 C 233kPa 5:18
136.4 C 230kPa 5:48
136.2 C 231kPa 6:18
136.2 C 231kPa 6:48
135.9 C 229kPa 7:18
136.4 C 233kPa 7:48
136.4 C 233kPa 7:48
VENTING 7:48
AIR DRYING 8:42
CYCLE COMPLETE 23:42

Digital Signature #
A79DAEFD7454BA02
