

-----  
Midmark M11 - v1.0.5  
Total Cycles: 559

Name: RITTER2  
Sterilizer ID: V2313228  
11/17/2021 01:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.4 F 0.0  
2:00 204.1 F 4.4  
4:00 210.1 F 10.6  
6:00 222.3 F 7.0  
8:00 237.5 F 13.1  
10:00 238.7 F 11.4  
12:00 239.3 F 10.6  
14:00 252.8 F 17.6  
16:00 264.7 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.5 F 29.9  
2:00 271.5 F 29.6  
3:00 272.4 F 29.9  
4:00 271.4 F 29.4  
  
Min 271.4 F 29.2  
Max 272.6 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.3 F 0.0  
5:00 218.3 F 0.0  
10:00 223.7 F 0.0  
15:00 221.6 F 0.0  
20:00 220.1 F 0.0  
25:00 220.9 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 2:29  
HEATING: 17:49  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:32  
CYCLE COMPLETE

11/17/2021 02:20 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_