
Midmark M11 - v1.0.0
Total Cycles: 3012

Name: RITTER 1
Sterilizer ID: V2178198
03/16/2023 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.9 F 0.0
2:00 187.4 F 3.1
4:00 194.2 F 7.2
6:00 213.5 F 12.6
8:00 225.3 F 7.5
10:00 237.7 F 12.9
12:00 238.1 F 10.4
14:00 235.9 F 8.6
16:00 250.5 F 15.7
18:00 262.5 F 22.5
20:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.3
2:00 272.1 F 28.7
3:00 271.5 F 28.5
4:00 272.6 F 29.3

Min 271.5 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 230.1 F 0.0
10:00 230.8 F 0.0
15:00 227.3 F 0.0
20:00 224.1 F 0.0
25:00 221.1 F 0.0
30:00 217.9 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 20:20
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:03
CYCLE COMPLETE

03/16/2023 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____