
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004953
9:39 17/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:57
STERILIZING 3:54
136.0 C 220kPa 3:54
136.3 C 222kPa 4:24
136.3 C 222kPa 4:54
136.4 C 221kPa 5:24
136.3 C 223kPa 5:54
136.1 C 218kPa 6:24
136.2 C 218kPa 6:54
136.4 C 221kPa 7:24
136.4 C 221kPa 7:24
VENTING 7:24
AIR DRYING 8:09
CYCLE COMPLETE 10:11

Digital Signature #
5B525B97E08A640E
