
Midmark M11 - v1.0.5
Total Cycles: 3391

Name: RITTER2
Sterilizer ID: V2313228
02/16/2024 12:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.9 F 0.0
2:00 146.1 F 1.4
4:00 160.4 F 4.3
6:00 183.1 F 8.1
8:00 201.4 F 12.5
10:00 214.2 F 6.3
12:00 228.0 F 11.0
14:00 229.9 F 7.9
16:00 241.0 F 12.3
18:00 244.1 F 12.9
20:00 256.3 F 19.3
22:00 267.4 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.5 F 30.0
2:00 271.4 F 29.3
3:00 272.7 F 29.7
4:00 271.7 F 29.7

Min 270.9 F 28.9
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 237.5 F 0.0
10:00 203.0 F 0.0
15:00 186.2 F 0.0
20:00 182.6 F 0.0
25:00 180.8 F 0.0
30:00 179.7 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:20
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:53
CYCLE COMPLETE

02/16/2024 01:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____