

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1286

Name: RITTER 1  
Sterilizer ID: V2178198  
10/21/2021 11:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.6 F 0.0  
2:00 187.1 F 3.0  
4:00 204.4 F 8.3  
6:00 224.2 F 8.6  
8:00 230.9 F 9.1  
10:00 240.0 F 10.1  
12:00 243.5 F 12.4  
14:00 242.0 F 11.2  
16:00 255.5 F 18.1  
18:00 267.3 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.3 F 28.7  
2:00 271.4 F 28.3  
3:00 272.3 F 28.8  
4:00 271.4 F 28.1  
  
Min 271.3 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.0 F 0.0  
5:00 219.3 F 0.0  
10:00 218.6 F 0.0  
15:00 227.6 F 0.0  
20:00 225.9 F 0.0  
25:00 228.3 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 2:31  
HEATING: 19:24  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:01  
CYCLE COMPLETE

10/21/2021 12:23 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_