

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3543

Name: RITTER2  
Sterilizer ID: V2313228  
04/18/2024 02:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 97.2 F 0.0  
2:00 97.3 F 0.0  
4:00 110.7 F 0.2  
6:00 139.5 F 2.2  
8:00 162.1 F 5.4  
10:00 185.1 F 9.1  
12:00 204.8 F 13.0  
14:00 218.4 F 7.1  
16:00 227.6 F 11.5  
18:00 230.7 F 8.3  
20:00 240.9 F 12.5  
22:00 242.9 F 12.4  
24:00 254.2 F 18.4  
26:00 264.9 F 24.9  
28:00 271.1 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.4  
1:00 272.3 F 30.2  
2:00 272.9 F 30.1  
3:00 271.9 F 29.6  
4:00 271.2 F 29.3

Min 271.0 F 29.2  
Max 273.1 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.5 F 0.0  
5:00 229.5 F 0.0  
10:00 223.3 F 0.0  
15:00 215.6 F 0.0  
20:00 216.3 F 0.0  
25:00 217.4 F 0.0  
30:00 216.9 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 28:04  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:06:34  
CYCLE COMPLETE

04/18/2024 03:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_