

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1935

Name: RITTER2  
Sterilizer ID: V2313228  
01/17/2023 09:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.2 F 0.0  
2:00 78.6 F 0.0  
4:00 111.5 F 0.0  
6:00 109.1 F 0.0  
8:00 160.4 F 0.1  
10:00 192.8 F 0.3  
12:00 204.8 F 3.9  
14:00 216.8 F 7.2  
16:00 226.8 F 11.0  
18:00 232.1 F 7.2  
20:00 240.8 F 12.3  
22:00 239.5 F 10.5  
24:00 237.0 F 9.3  
26:00 251.4 F 16.3  
28:00 263.1 F 23.1  
30:00 272.3 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.6 F 29.3  
2:00 271.5 F 28.7  
3:00 272.6 F 29.4  
4:00 271.7 F 28.7

Min 271.4 F 28.6  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.7 F 0.0  
5:00 213.8 F 0.0  
10:00 225.4 F 0.0  
15:00 225.5 F 0.0  
20:00 229.5 F 0.0  
25:00 229.7 F 0.0  
30:00 226.4 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:13  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:08:09  
CYCLE COMPLETE

01/17/2023 10:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_