

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1208

Name: RITTER 1  
Sterilizer ID: V2178198  
09/27/2021 01:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 99.7 F 0.0  
2:00 102.9 F 0.0  
4:00 114.9 F 0.1  
6:00 157.2 F 0.2  
8:00 192.1 F 1.5  
10:00 208.7 F 5.1  
12:00 221.0 F 8.7  
14:00 229.2 F 12.6  
16:00 231.0 F 8.3  
18:00 242.7 F 13.4  
20:00 240.8 F 11.2  
22:00 237.6 F 9.5  
24:00 251.6 F 16.0  
26:00 263.1 F 22.6  
28:00 272.3 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.4 F 28.8  
2:00 271.4 F 28.0  
3:00 272.5 F 29.0  
4:00 272.0 F 28.4

Min 271.3 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.9 F 0.0  
5:00 213.0 F 0.0  
10:00 216.5 F 0.0  
15:00 214.7 F 0.0  
20:00 214.2 F 0.0  
25:00 216.0 F 0.0  
30:00 213.6 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 28:16  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:06:15  
CYCLE COMPLETE

09/27/2021 02:54 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01