
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002507
9:05 11/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:39
PRESSURIZING 3:41
STERILIZING 4:48
136.0 C 221kPa 4:48
136.2 C 223kPa 5:18
136.4 C 224kPa 5:48
136.4 C 224kPa 6:18
136.4 C 224kPa 6:48
136.0 C 222kPa 7:18
136.3 C 223kPa 7:48
136.3 C 224kPa 8:18
136.3 C 224kPa 8:18
VENTING 8:18
AIR DRYING 9:09
CYCLE COMPLETE 15:53

Digital Signature #
40B798D6A2E0B008
