
Midmark M11 - v1.0.5
Total Cycles: 3518

Name: RITTER2
Sterilizer ID: V2313228
04/11/2024 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.9 F 0.0
2:00 141.8 F 2.2
4:00 163.2 F 5.3
6:00 189.5 F 9.1
8:00 209.9 F 13.5
10:00 216.5 F 7.6
12:00 229.4 F 12.4
14:00 231.7 F 8.9
16:00 239.4 F 10.9
18:00 244.1 F 13.2
20:00 255.9 F 19.5
22:00 266.8 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.5
1:00 272.9 F 30.1
2:00 271.5 F 29.4
3:00 271.8 F 30.1
4:00 272.7 F 30.0

Min 271.1 F 29.3
Max 273.2 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 201.4 F 0.0
10:00 201.6 F 0.0
15:00 194.0 F 0.0
20:00 192.7 F 0.0
25:00 198.3 F 0.0
30:00 206.8 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:34
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:02:10
CYCLE COMPLETE

04/11/2024 02:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____