
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001397
9:59 25/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:36
STERILIZING 3:43
136.4 C 223kPa 3:43
136.2 C 224kPa 4:13
136.4 C 223kPa 4:43
136.3 C 223kPa 5:13
136.3 C 225kPa 5:43
136.3 C 223kPa 6:13
136.3 C 224kPa 6:43
136.3 C 224kPa 7:13
136.3 C 224kPa 7:13
VENTING 7:13
AIR DRYING 8:03
CYCLE COMPLETE 8:08

Digital Signature #
0D57C6171293B666
