
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007855
7:42 11/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:53
STERILIZING 4:02
136.4 C 221kPa 4:02
136.3 C 222kPa 4:32
136.3 C 219kPa 5:02
136.3 C 221kPa 5:32
136.5 C 221kPa 6:02
136.4 C 221kPa 6:32
136.4 C 223kPa 7:02
136.4 C 222kPa 7:32
136.4 C 222kPa 7:32
VENTING 7:32
AIR DRYING 8:23
CYCLE COMPLETE 38:23

Digital Signature #
1868F17AB0ACA890
