
Midmark M11 - v1.0.0
Total Cycles: 1964

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2022 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.5 F 0.0
2:00 209.4 F 5.2
4:00 214.9 F 10.8
6:00 228.0 F 7.1
8:00 239.7 F 13.0
10:00 239.4 F 10.9
12:00 237.8 F 9.6
14:00 252.9 F 16.9
16:00 265.2 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.4 F 29.1
2:00 271.5 F 28.4
3:00 272.5 F 29.2
4:00 271.7 F 28.5

Min 271.4 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 221.4 F 0.0
10:00 242.5 F 0.0
15:00 246.0 F 0.0
20:00 240.4 F 0.0
25:00 225.9 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 17:43
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:07
CYCLE COMPLETE

05/27/2022 11:37 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____