

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000286  
8:46 09/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:23  
PRESSURIZING 2:38  
STERILIZING 4:02  
136.0 C 229kPa 4:02  
136.4 C 230kPa 4:32  
136.1 C 230kPa 5:02  
136.4 C 233kPa 5:32  
136.4 C 231kPa 6:02  
136.1 C 230kPa 6:32  
136.4 C 233kPa 7:02  
136.4 C 232kPa 7:32  
136.4 C 232kPa 7:32  
VENTING 7:32  
AIR DRYING 8:26  
CYCLE COMPLETE 8:31  
-----

Digital Signature #  
2D86E4A81A7231C8  
-----