
Midmark M11 - v1.0.0
Total Cycles: 5447

Name: RITTER 1
Sterilizer ID: V2178198
03/26/2025 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.5 F 0.0
2:00 130.0 F 1.7
4:00 143.2 F 4.6
6:00 164.9 F 8.0
8:00 187.3 F 11.8
10:00 207.5 F 3.2
12:00 219.5 F 7.2
14:00 229.9 F 11.6
16:00 230.6 F 7.3
18:00 243.0 F 12.6
20:00 240.5 F 10.7
22:00 252.5 F 16.6
24:00 263.1 F 22.8
26:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.2 F 29.2
2:00 272.0 F 28.7
3:00 271.2 F 28.4
4:00 272.0 F 29.2

Min 271.2 F 28.3
Max 272.3 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 199.3 F 0.0
10:00 210.1 F 0.0
15:00 199.6 F 0.0
20:00 195.0 F 0.0
25:00 193.4 F 0.0
30:00 194.2 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 26:23
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:05:22
CYCLE COMPLETE

03/26/2025 11:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____