
Midmark M11 - v1.0.0
Total Cycles: 5205

Name: RITTER 1
Sterilizer ID: V2178198
12/18/2024 05:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.6 F 0.0
2:00 175.5 F 0.4
4:00 199.8 F 3.9
6:00 215.7 F 8.3
8:00 227.7 F 13.1
10:00 231.5 F 8.5
12:00 241.9 F 12.3
14:00 242.6 F 11.8
16:00 241.0 F 10.7
18:00 254.3 F 17.3
20:00 265.6 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.6 F 28.7
2:00 271.7 F 28.1
3:00 271.8 F 28.5
4:00 272.5 F 28.7

Min 271.3 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.0 F 0.0
5:00 204.3 F 0.0
10:00 201.1 F 0.0
15:00 194.1 F 0.0
20:00 191.3 F 0.0
25:00 191.9 F 0.0
30:00 197.0 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:46
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:00:36
CYCLE COMPLETE

12/18/2024 06:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____