
Midmark M11 - v1.0.5
Total Cycles: 890

Name: RITTER2
Sterilizer ID: V2313228
03/21/2022 12:12 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.6 F 0.0
2:00 127.2 F 0.1
4:00 181.2 F 0.4
6:00 185.4 F 2.9
8:00 206.8 F 9.3
10:00 225.1 F 7.1
12:00 241.8 F 13.5
14:00 243.6 F 12.6
16:00 244.0 F 12.4
18:00 258.8 F 20.4
20:00 271.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.0 F 28.6
2:00 272.7 F 29.1
3:00 272.6 F 29.4

Min 271.4 F 28.4
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 220.4 F 0.0
5:00 228.9 F 0.0
10:00 220.7 F 0.0
15:00 216.1 F 0.0
20:00 213.5 F 0.0
25:00 215.6 F 0.0
30:00 212.8 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 20:34
STERILIZING: 3:00
VENTING: 2:46
DRYING: 30:00
TOTAL CYCLE: 00:58:00
CYCLE COMPLETE

03/21/2022 01:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____