
Midmark M11 - v1.0.5
Total Cycles: 4142

Name: RITTER2
Sterilizer ID: V2313228
11/20/2024 02:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.6 F 0.0
2:00 145.1 F 0.1
4:00 171.0 F 2.3
6:00 190.1 F 6.1
8:00 206.6 F 9.9
10:00 221.0 F 8.4
12:00 228.2 F 8.9
14:00 237.7 F 12.1
16:00 240.0 F 11.6
18:00 239.2 F 10.5
20:00 251.7 F 16.7
22:00 262.7 F 23.0
24:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.7
2:00 271.8 F 29.1
3:00 271.4 F 29.3
4:00 272.5 F 29.7

Min 271.1 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 228.3 F 0.0
10:00 223.5 F 0.0
15:00 216.4 F 0.0
20:00 213.4 F 0.0
25:00 208.3 F 0.0
30:00 206.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:25
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:09
CYCLE COMPLETE

11/20/2024 03:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____