
Midmark M11 - v1.0.0
Total Cycles: 5453

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2025 08:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 93.2 F 0.3
4:00 86.6 F 1.7
6:00 125.7 F 4.4
8:00 156.1 F 7.1
10:00 179.9 F 10.5
12:00 187.5 F 5.1
14:00 207.1 F 4.9
16:00 219.0 F 8.5
18:00 229.9 F 12.4
20:00 229.9 F 7.4
22:00 241.4 F 12.2
24:00 238.2 F 9.6
26:00 250.7 F 15.6
28:00 261.6 F 21.8
30:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.7 F 28.8
2:00 272.1 F 28.7
3:00 272.4 F 28.6
4:00 272.2 F 28.8

Min 271.6 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 178.4 F 0.0
10:00 189.6 F 0.0
15:00 197.0 F 0.0
20:00 197.3 F 0.0
25:00 198.8 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 30:31
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:08:43
CYCLE COMPLETE

03/27/2025 10:05 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____