

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3919

Name: RITTER2  
Sterilizer ID: V2313228  
09/09/2024 04:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 98.5 F 0.0  
2:00 100.1 F 0.0  
4:00 105.7 F 0.2  
6:00 137.6 F 2.3  
8:00 168.7 F 5.5  
10:00 202.5 F 9.2  
12:00 214.7 F 13.2  
14:00 224.0 F 8.3  
16:00 235.2 F 13.2  
18:00 235.8 F 10.6  
20:00 234.8 F 8.8  
22:00 248.6 F 15.4  
24:00 260.1 F 21.9  
26:00 270.6 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 30.0  
1:00 272.4 F 29.9  
2:00 272.6 F 30.3  
3:00 272.5 F 30.6  
4:00 271.9 F 30.1

Min 271.8 F 29.9  
Max 272.8 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
5:00 207.1 F 0.0  
10:00 228.4 F 0.0  
15:00 228.8 F 0.0  
20:00 227.5 F 0.0  
25:00 222.1 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 27:28  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:05:57  
CYCLE COMPLETE

09/09/2024 05:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_