
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006242
7:36 06/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:56
STERILIZING 3:50
136.0 C 220kPa 3:50
136.4 C 222kPa 4:20
136.4 C 222kPa 4:50
136.4 C 223kPa 5:20
136.2 C 221kPa 5:50
136.2 C 222kPa 6:20
136.1 C 221kPa 6:50
136.4 C 220kPa 7:20
136.4 C 220kPa 7:20
VENTING 7:20
AIR DRYING 8:11
CYCLE COMPLETE 11:40

Digital Signature #
DD0C0026E4AF1709
