

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3786

Name: RITTER 1  
Sterilizer ID: V2178198  
09/18/2023 09:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.1 F 0.0  
2:00 160.3 F 0.0  
4:00 168.0 F 1.4  
6:00 191.1 F 4.5  
8:00 210.1 F 8.2  
10:00 224.2 F 12.5  
12:00 229.8 F 7.4  
14:00 239.8 F 12.3  
16:00 237.5 F 9.8  
18:00 238.2 F 9.2  
20:00 248.4 F 14.4  
22:00 259.7 F 20.6  
24:00 270.0 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.5 F 28.9  
2:00 271.3 F 28.2  
3:00 272.2 F 29.0  
4:00 272.0 F 28.6

Min 271.1 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 172.3 F 0.0  
5:00 204.7 F 0.0  
10:00 224.9 F 0.0  
15:00 228.0 F 0.0  
20:00 235.8 F 0.0  
25:00 236.7 F 0.0  
30:00 222.5 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 24:58  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:03:36  
CYCLE COMPLETE

09/18/2023 10:29 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_